

MINTec-SINAGAMA
AGENSI NUKLEAR MALAYSIA
(formerly known as Malaysian Institute For Nuclear Technology Research - Mint)
BANGI, 43000 KAJANG, SELANGOR DARUL EHSAN.
Tel : 03-89112000 Fax.: 03-89201099 e-mel: sgama@nuclearmalaysia.gov.my

Ruj. Tuan : DO NO: OSSCARE-0026/21
Your Ref. :
Ruj. Kami : QC/KP/21 (S:0674/21)
Our Ref. :

KEPUTUSAN ANALISIS PENYINARAN
Results of Irradiation Analysis

Nama dan Alamat Pelanggan
Customer Name and Address

OSS CARE MANUFACTURING SDN. BHD.
NO. 26, JALAN INDUSTRI IMPIAN 1,
TAMAN INDUSTRI IMPIAN,
14000 BUKIT MERTAJAM,
PULAU PINANG.
(ATTN: MR. BALAN)

Keterangan Bahan
Product Description

PRODUCT NAME	PRODUCT CODE	BATCH
CLEANROOM WIPER	ISO-0909S	536683/200921001

Ukuran Bungkusan/Kuantiti
Product Dimension/Quantity

: 390 CTNS
(22.0 m³)

Dose Permohonan
Requested Irradiation Dose

: Dose Min = 25 kGy
Dose Max = 40 kGy

Tarikh Penyinaran
Irradiation Date

: 10/11/2021

Keputusan
Results


: Dose Min = 26.8 kGy
Dose Max = 29.3 kGy

Catatan
Remarks

:

Disahkan :-
(Approved by)

Tarikh :- 10/11/2021
Date



AHSANULKHALIQIN ABDUL WAHAB
Sinagama
Manager
Irradiation Facilities Division
Malaysian Nuclear Agency